

**Effect of Microtox<sup>®</sup> Reagent Reconstitution Age  
on the Variability of Analytical Results  
from the Microtox<sup>®</sup> Assay**

**Final Report**

to

Western Canada Microtox Users Committee

by

**I.D. Gaudet**

Microbiology Research & Development,  
Alberta Research Council

## EXECUTIVE SUMMARY

This report presents the results of a study into the effect of reconstitution age of Microtox® reagent on the sensitivity and reproducibility of the Microtox assay. Additionally, the study dealt with the effect of lot to lot variation on the sensitivity and reproducibility of the Microtox assay. Four different samples (zinc sulphate, copper sulphate, phenol, and sodium dodecyl sulphate) were evaluated at seven different reconstitution times from 0.5 to 24 hours (0.5, 0.75, 1.0, 2.0, 4.0, 8.0, 24.0) using the same vial of reagent. Each sample was analyzed in triplicate with three different lots of reagent. Statistical analysis of results focused on 15 minute EC<sub>50</sub> data only.

Several conclusions were drawn from this study. Firstly, it was found that there were differences in toxicity over time with different compounds. With the four compounds tested, all three possibilities of an increase, a decrease, and no change in toxicity with reagent age were observed. Secondly, it was observed that due to the increase or decrease in toxicity with aged reagent, data obtained from reagent which is older than 4 hours should be regarded cautiously, if at all. Finally, it was observed that there is lot to lot variation in toxicity results for some of the compounds tested.

Several recommendations were made based on the results of this study.

1. Each lot of reagent should have quality control testing performed using standard reference toxicant(s).
2. Bench recording sheets or computer recording programs should include the time of reconstitution of reagent as well as the start time of the assay.
3. Microtox® test reagent should generally not be used for testing after 4 hours from time of reconstitution. Any assay performed after this time may yield unpredictable results due to reagent age and not the toxicity of the sample being tested. Results obtained from reagent used after 4 hours needs to be evaluated cautiously, with the use of an appropriate control toxicant run in parallel.
4. WCMUC adopt these points within their standard operating procedure and as part of their ongoing inter-laboratory quality control program.

---

**TABLE OF CONTENTS**

<b>Section</b>	<b>Page</b>
<b>A. INTRODUCTION</b>	1
<b>B. OBJECTIVES AND SCOPE</b>	1
<b>1. Objectives</b>	1
<b>2. Scope</b>	2
<b>C. PROCEDURES</b>	2
<b>1. Microtox® Analysis</b>	2
<b>2. Statistical Analysis</b>	2
<b>D. RESULTS AND DISCUSSION</b>	3
<b>1. Zinc</b>	4
<b>2. Copper</b>	5
<b>3. Phenol</b>	5
<b>4. Sodium Dodecyl Sulphate</b>	6
<b>E. CONCLUSIONS</b>	6
<b>F. RECOMMENDATIONS</b>	7
<b>G. ACKNOWLEDGEMENTS</b>	8
<b>H. REFERENCES</b>	8
<b>I. APPENDICES</b>	10
<b>1. AZUR Quality Control Data</b>	11
Phenol Quality Control Chart	11
Zinc Sulphate Quality Control Chart	12
Certificate of Performance, Lot ACV009-2	13
Certificate of Performance, Lot ACV009-3	14
Certificate of Performance, Lot ACV010-3	15
<b>2. Tables and Figures</b>	16
a) Zinc	16
Table 1. Results of Zinc Sulphate Assays	16
Table 2. Zinc Sulphate EC <sub>50</sub> 15 minute Data Summary	17
Figure 1. Zinc Sulphate EC <sub>50</sub> 15 Minute Data	18
Figure 2. Zinc Sulphate EC <sub>50</sub> 15 Minute Data (24 hour data removed)	19
Table 3. Zinc Sulphate Differences Among Mean EC <sub>50</sub> Values	20
Table 4. Zinc Sulphate Differences Among Means, Reagent Lots EC <sub>50</sub> Values	20
b) Copper	21
Table 5. Results of Copper Sulphate Assays	21
Table 6. Copper Sulphate EC <sub>50</sub> 15 minute Data Summary	22
Figure 3. Copper Sulphate EC <sub>50</sub> 15 Minute Data	23
Figure 4. Copper Sulphate EC <sub>50</sub> 15 Minute Data (24 hour data removed)	24
Figure 5. Copper Sulphate EC <sub>50</sub> 15 Minute Data (8 and 24 hour data removed)	25

**Western Canada Microtox® Users Committee – Final Report**

---

Table 7.	Copper Sulphate Differences Among Mean EC <sub>50</sub> Values	26
Table 8.	Copper Sulphate Differences Among Means, Reagent Lots EC <sub>50</sub> Values	26
c) Phenol		27
Table 9.	Results of Phenol Assays	27
Table 10.	Phenol EC <sub>50</sub> 5 minute Data Summary	28
Figure 6.	Phenol EC <sub>50</sub> 5 minute Data	29
Table 11.	Phenol EC <sub>50</sub> 15 minute Data Summary	30
Figure 7.	Phenol EC <sub>50</sub> 15 minute Data	31
Table 12.	Phenol Differences Among Mean EC <sub>50</sub> Values	32
Table 13.	Phenol Differences Among Means, Reagent Lots EC <sub>50</sub> Values	32
d) Sodium Dodecyl Sulphate		33
Table 14.	Results of Sodium Dodecyl Sulphate Assays	33
Table 15.	Sodium Dodecyl Sulphate EC <sub>50</sub> 15 minute Data Summary	34
Figure 8.	Sodium Dodecyl Sulphate EC <sub>50</sub> 15 minute Data	35
Figure 9.	Sodium Dodecyl Sulphate EC <sub>50</sub> 15 minute Data (24 hour data removed)	36
Table 16.	Sodium Dodecyl Sulphate Differences Among Mean EC <sub>50</sub> Values	37
Table 17.	Sodium Dodecyl Sulphate Differences Among Means, Reagent Lots EC <sub>50</sub> Values	37

## A. INTRODUCTION

The Microtox® basic test assay has been used to determine the toxicity of such diverse samples as oil sands tailings water (MacKinnon and Retallack, 1982; Baddaloo, 1986), landfill leachates (Cameron *et al.*, 1982; Baker, 1985;), pulp and paper effluent (Blaise *et al.*, 1987), drilling fluids (Stroscher, 1984) and several industrial effluents (Bulich *et al.*, 1981; Vasseur *et al.*, 1984; Coleman and Qureshi, 1985). It has been generally accepted that there is no universal bioassay which can be used in every situation for toxicity testing and that a battery of screening tests should be used. Due to the ease of use, low cost, sensitivity, and reproducibility, the Microtox test system has often been used as a pre-screening tool.

Several refinements to the Microtox basic test assay have been made since its inception. One of the first was the use of colour correction to determine the toxicity of highly coloured samples (Beckman, 1982). In a study on the refinement of the Microtox assay, Qureshi *et al.* (1984) made several recommendations for the assay. First, the sample pH should be considered part of the assay since adjustments cause changes within the sample. Secondly, the 15 minute EC<sub>50</sub> should be accepted as the standard endpoint for assaying samples, particularly those containing metallic ions. These two recommendations have found general acceptance. The third major recommendation, that freshly reconstituted reagent be used for testing, does not seem to have been generally accepted. This may be due to the expense of the reagent, time constraints within a testing laboratory, or other unknown factors. A recent publication lists the following bacterial factors as causes of variability of the toxicity test results: age of freeze-dried bacteria, bacterial storage and conservation, reconstitution procedure, age of the reconstituted reagent, and temperature equilibration time (Ribo, 1997). However, Ribo did not present any results which would address these issues.

To address the question of the effect of reconstitution age, this study investigated the effect that the reconstitution age of the Microtox® test reagent has on the sensitivity of the Microtox assay. This study investigated the effect of reconstitution age of the Microtox® test reagent on the sensitivity of the Microtox assay when tested against an array of compounds and the effect that lot to lot variation may have on the effect of reconstitution age of the Microtox® test reagent.

## B. OBJECTIVES & SCOPE

### 1. Objectives.

The objectives of this study were to:

1. investigate the effect of reconstitution age of Microtox® test reagent on the results obtained with the assay,

2. investigate the effect of lot to lot variation of Microtox® test reagent, and
3. provide a possible recommendation as to the maximum time after reconstitution for use of the Microtox® test reagent.

## **2. Scope**

The scope of the work plan for each of the components was as follows. This study evaluated four different samples at seven different reconstitution times between 0.5 to 24 hours (0.5, 0.75, 1.0, 2.0, 4.0, 8.0, 24.0) using the same vial of reagent. The samples consisted of two metallic salts, copper sulphate and zinc sulphate and two organics, phenol and sodium dodecyl sulphate (SDS) which is also called sodium lauryl sulphate. Each sample was analyzed in triplicate. This series was repeated with three different reagent lots, AVC009-2, AVC009-3, and AVC010-3. Quality control data of each lot of reagent as well as quality control charts for phenol and zinc sulphate as supplied by AZUR Environmental are presented in Appendix 1. Statistical analysis of results focused on 15 minute EC<sub>50</sub> data only since this is the time point most commonly used for reporting of Microtox testing results.

## **C. PROCEDURES**

### **1. Microtox® Analysis**

The analytical procedure for using the Model 500 analyzer for Microtox® testing has been described previously (Gaudet, 1994). The assays were performed using the standard 82% assay (AEC variation) (Gaudet, 1994). Timing for the assays was determined as follows. Reagent was reconstituted, the clock was started and reagent was allowed to sit in the reagent well for 15 minutes. Reagent was then diluted into the testing wells and allowed to equilibrate for an additional 15 minutes. Initial light readings were taken and, at 30 minutes from the time of reconstitution, the toxicant was added to the testing wells. Time was determined as the total elapsed time from reconstitution to the addition of the toxicant, allowing for temperature equilibration. To run additional assays from the same reagent vial, the reagent was again diluted into testing wells and allowed to equilibrate to run analysis at the additional time points. Reagent was added to diluent 15 minutes prior to running an assay to allow for temperature equilibration.

### **2. Statistical Analysis**

For the statistical analysis, two questions were asked. Firstly, “is there a difference between the toxicity values at the various time points?” In order to determine this statistical comparison, a standard time of 1 hour was chosen and

all other results were compared to this standard time. The second question was “is there a difference between the reagent lots at each time point?”

**Differences in EC<sub>50</sub> values at various time points against the time point ‘1 hour’.** The statistical model for each reagent was the One-way Analysis of Variances. The general form of the model is given by

$$[EC_{50}]_{ijk} = \mu + \beta_{ij} * [Time(Hour)]_{ijk} + \varepsilon_{ijk} ,$$

where

$\mu$  = overall mean,

$\beta_{ij}$ ,  $j = 0.5, 0.75, 1, 2, 4, 8, 24$  = main effect of Time for each

$i = ACV009-2, ACV009-3, ACV010-3,$

$\varepsilon_{ijk} \sim iid N(0, \sigma^2)$  with  $k = 1, 2, 3.$

For each reagent  $i$ , the mean values of EC<sub>50</sub> of various Time points were compared against the mean EC<sub>50</sub> of normal time ‘1 hour’ using Dunnett’s comparison method in GLM procedure in SAS/STAT software (SAS Institute Inc. 1990a, 1990b, 1994).

**Differences of EC<sub>50</sub> values between reagent lots at each time point.** The objective of this statistical analysis was to examine whether EC<sub>50</sub> values of different reagents are significantly different at each time point. The statistical model for this analysis was the One-way Analysis of Covariance. The general form of the model is given by

$$[EC_{50}]_{ijk} = \mu + \alpha_i + \beta_j * [Time(Hour)]_{ijk} + (\alpha\beta)_{ij} + \varepsilon_{ijk} ,$$

where

$\mu$  = overall mean,

$\alpha_i$ ,  $i = ACV009-2, ACV009-3, ACV010-3$  = main effect of Reagent,

$\beta_j$ ,  $j = 0.5, 0.75, 1, 2, 4, 8, 24$  = main effect of Time,

$(\alpha\beta)_{ij}$  = interaction between Reagent and Time,

$\varepsilon_{ijk} \sim iid N(0, \sigma^2)$  with  $k = 1, 2, 3.$

After the model was fitted using GLM procedure in SAS/STAT software from SAS institute (SAS Institute Inc. 1990a, 1990b, 1994), the adjusted means of EC<sub>50</sub> for each reagent within each time point were compared against each other using Tukey’s method.

## D. RESULTS AND DISCUSSION

The results of Microtox testing of each sample will be reported and discussed separately. However, for better understanding, general results are described first. Results of each series of analyses is presented in Appendix 2 – Tables and Figures. The first table in each series lists the EC<sub>20</sub> and EC<sub>50</sub> at 5 and 15 minutes for all reagent lots, all replicates, and all reconstitution times tested. The second table

contains information that recalculates the 15 minute EC<sub>50</sub> results and gives arithmetic mean values for each reconstitution time and reagent lot as well as overall means for each. For phenol, the 5 minute data was also recalculated, since AZUR Environmental quality control data for phenol is given for 5 minute assays. Following the reworked data is a graphical representation of these results. Since some compounds showed a sharp increase or decrease in toxicity at 24 hours and occasionally at 8 hours, there are additional graphs presented with these results removed for zinc (24 hour data removed), copper (24 hour data removed, and 8 and 24 hour data removed), and SDS (24 hour data removed). Next is the statistical summary for each compound indicating the arithmetic means, standard error for each reagent lot and each reconstitution time for the EC<sub>50</sub> data only. Finally, each section concludes with a table which compares the EC<sub>50</sub> values at various reconstitution times for different lots to see if there is a difference between the lots. Results presented are adjusted means and standard error of those means and a multiple comparison for each time point between lots.

## **1. Zinc**

Table 1 summarizes the EC<sub>50</sub> and EC<sub>20</sub> data at 5 and 15 minutes for all zinc sulphate assays. EC<sub>50</sub> 15 minute results are summarized in Table 2. A graphical representation of 15 minute EC<sub>50</sub> results is presented in Figures 1 and 2. Statistical results are presented in Tables 3 and 4.

The overall mean EC<sub>50</sub> 15 minutes for zinc sulphate was 1.291 mg/L Zn<sup>+2</sup> for all reconstitution times (Figure 1). However, a sharp increase in toxicity was observed at 24 h and these assays had an overall EC<sub>50</sub> of 0.518 mg/L Zn<sup>+2</sup>. When the 24 hour data were removed, the mean EC<sub>50</sub> was 1.420 mg/L Zn<sup>+2</sup> (Figure 2).

AZUR certificates of performance list zinc sulphate EC<sub>50</sub> 15 minutes at 0.6 to 2.2 mg/L Zn<sup>+2</sup> (Appendix 1). All assays, except for the majority of those performed at 24 hours, fall into these general specifications. In addition, most assays results were similar to the AZUR quality assurance results which list the zinc sulphate EC<sub>50</sub> 15 minutes (mg/l Zn<sup>+2</sup>) at 1.3 for lot ACV009-2, 1.0 for lot ACV009-3, and 1.2 for lot ACV010-3. EC<sub>50</sub> results for this study ranged from a low of 0.835 to a high of 2.123 (Table 2).

Statistical analysis of EC<sub>50</sub> results compared to the chosen standard time of 1 hour found that for all lots, the 24 hour data was significantly different (Table 3). Therefore, the rejection of 24 hour data in Figure 2 was statistically reinforced. In addition, for lot ACV009-2 the EC<sub>50</sub> results for 0.5, 0.75 and 8 hours were significantly different from the 1 hour results (Table 3). However,

when adjusted means of EC<sub>50</sub> results were compared, there was no difference between the 3 reagent lots at all time points tested (Table 4).

## **2. Copper**

Table 5 summarizes the EC<sub>50</sub> and EC<sub>20</sub> data at 5 and 15 minutes for all copper sulphate assays. EC<sub>50</sub> 15 minute results are summarized in Table 6. A graphical representation of 15 minute EC<sub>50</sub> results is presented in Figures 3, 4 and 5. Statistical results are presented in Tables 7 and 8.

Copper sulphate had an overall mean EC<sub>50</sub> at 15 minutes of 0.308 mg/L Cu<sup>+2</sup> for all reconstitution times (Figure 3). However, a sharp decrease in toxicity was observed at 24 h (overall EC<sub>50</sub> of 0.461 mg/L Cu<sup>+2</sup>). When the 24 hour data were removed, the mean EC<sub>50</sub> was 0.282 mg/L Cu<sup>+2</sup> (Figure 4). The increase in toxicity was still observed at 8 hours (overall EC<sub>50</sub> of 0.417 mg/L Cu<sup>+2</sup>). When the 24 and 8 hour data were removed, the mean EC<sub>50</sub> was 0.256 mg/L Cu<sup>+2</sup> (Figure 5) ranging from a low of 0.156 to a high of 0.406.

Statistical analysis of EC<sub>50</sub> results compared to the chosen standard time of 1 hour found that, for all lots, the 8 and 24 hour data were significantly different (Table 7). Thus, the rejection of these time points as presented in Figure 5 was statistically reinforced. In addition, the comparisons of adjusted means of EC<sub>50</sub> results found that there was a difference between the 3 lots at the 4 hour and 8 hour time points (Table 8).

## **3. Phenol**

Table 9 summarizes the EC<sub>50</sub> and EC<sub>20</sub> data at 5 and 15 minutes for all phenol assays. EC<sub>50</sub> 5 minute results are summarized in Table 10. A graphical representation of 5 minute EC<sub>50</sub> results is presented in Figure 6. EC<sub>50</sub> 15 minute results are summarized in Table 11. A graphical representation of 15 minute EC<sub>50</sub> results is presented in Figure 7. Statistical results are presented in Tables 12 and 13.

Phenol had an overall mean EC<sub>50</sub> 5 minutes of 20.199 mg/L for all reconstitution times (Table 10) ranging from a low of 15.916 to a high of 25.228. AZUR quality control data lists phenol EC<sub>50</sub> 5 minutes at 13 to 26 mg/L. All assays at all time points fit into this specification, however, all are well above the AZUR quality assurance results for each lot which list the phenol EC<sub>50</sub> 5 minutes at 14.7 for lot ACV009-2, 14.3 for lot ACV009-3, and 13.3 for lot ACV010-3.

The EC<sub>50</sub> 15 minutes had an overall mean of 21.366 and ranged from a low of 16.601 to a high of 27.061 (Table 11). The graphical plot of these results (Figure 7) shows no particular trend in toxicity values, however, there does appear to be lot to lot variation with results from lot ACV009-2 being consistently high and results from lot ACV010-3 being consistently low. Statistical analysis of EC<sub>50</sub> results compared to the chosen standard time of 1 hour found that significantly different results were observed at 2 and 24 hours for reagent lot AVC009-2 and at 0.75 and 2 hours for reagent lot ACV010-3 (Table 12). In addition, adjusted means of EC<sub>50</sub> results comparison finds that there was a difference between the 3 lots at all time points except for 24 hours (Table 13). This reaffirms the observation of high/low distribution with particular reagent lots seen in Figure 7.

#### **4. Sodium Dodecyl Sulphate**

Table 14 summarizes the EC<sub>50</sub> and EC<sub>20</sub> data at 5 and 15 minutes for all sodium dodecyl sulphate assays. EC<sub>50</sub> 15 minute results are summarized in Table 15. A graphical representation of 15 minute EC<sub>50</sub> results is presented in Figures 8 and 9. Statistical results are presented in Tables 16 and 17.

SDS had an overall mean EC<sub>50</sub> at 15 minutes of 0.893 mg/L for all reconstitution times (Figure 8). However, a sharp decrease in toxicity was observed at 24 h (overall EC<sub>50</sub> of 1.708 mg/L). When the 24 hour data were removed, the mean EC<sub>50</sub> was 0.757 mg/L (Figure 9) ranging from a low of 0.503 to a high of 1.003 (Table 15).

Statistical analysis of EC<sub>50</sub> results compared to the chosen standard time of 1 hour found that for all lots, the EC<sub>50</sub> at 24 was significantly different from that at 1 hour (Table 16). Therefore, the rejection of 24 hour data as presented in Figure 9 is statistically reinforced. In addition, adjusted means of EC<sub>50</sub> comparison, finds that there was a difference between the 3 lots at 8 and 24 hours (Table 17).

### **E. CONCLUSIONS**

The first and most obvious conclusion to be drawn from this testing was that there were differences in toxicity over time for the compounds tested. Zinc exhibited a sharp increase in toxicity at 24 hours, however EC<sub>50</sub> results were relatively stable prior to that time. EC<sub>50</sub> ranged from a low of 0.835 to a high of 2.123 with the mean value excluding 24 hour data being 1.420. SDS also exhibited an increase in toxicity at 24 hours. Results prior to this time point were fairly stable as well. EC<sub>50</sub> ranged from a low of 0.503 to a high of 1.003 with the mean value excluding 24

hour data being 0.757. Copper, on the other hand, exhibited a decrease in toxicity starting at 4 hours which continued at the 8 and 24 hour testing points. Prior to this, however, toxicity results were very stable. EC<sub>50</sub> ranged from a low of 0.156 to a high of 0.406 with the mean value excluding 8 and 24 hour data being 0.256. Phenol produced no toxicity trend over time, however, throughout the 24 hour period, the results were erratic. EC<sub>50</sub> ranged from a low of 16.601 to a high of 27.061 with the mean value being 21.366. Thus, based on the results of all samples, a prediction of change in toxicity of an unknown compound or sample with time cannot be made. With the four compounds tested, all three possibilities of increase, decrease, or no change in toxicity were observed.

The second conclusion was that Microtox® test reagent should definitely not be used at 24 hours, and probably not at 8 hours. Except for phenol, the 24 hour data were rejected for all compounds tested and the 8 hour data were rejected for copper as well. Thus, data obtained from reagent which is older than 4 hours should be regarded cautiously, if at all.

When statistical analysis of data between lots was performed, it was found that there was variation in toxicity results between lots for copper, phenol, and SDS. With phenol, for example, the results for lot ACV009-2 are consistently higher than those for the other two lots except at 24 hours (Figure 7). Phenol shows a definite trend for lot to lot variation with EC<sub>50</sub> results being lowest for lot ACV010-3 and highest for lot ACV009-2. Statistical analysis for lot to lot variation for phenol indicated there was variation at all time points except for 24 hours (Table 13).

## **F. RECOMMENDATIONS**

There are several recommendations which arise from this study. Many of these are part of general practice of most laboratories performing the Microtox test. These recommendations are as follows.

1. Each lot of reagent should have quality control testing performed using standard reference toxicant(s). This long term testing would then indicate lot to lot variation and might indicate if there is a problem with a new lot of reagent. Testing should be performed on each new lot of reagent and at regular intervals during the shelf life of that lot.
2. Bench recording sheets or computer recording programs should include the time of reconstitution of reagent as well as the start time of the assay. In this manner, potential problems in interpretation of data caused by use of aged reagent can be avoided.
3. Microtox® test reagent should generally not be used for testing after 4 hours from time of reconstitution. Any assay performed after this time may yield unpredictable results due to reagent age and not toxicity of the sample being

tested. Results obtained from reagent used after 4 hours needs to be evaluated cautiously, perhaps with the use of an appropriate control toxicant run in parallel.

4. It is further recommended that WCMUC adopt these points within their standard operating procedure and as part of their ongoing inter-laboratory quality control program.

## **G. ACKNOWLEDGEMENTS**

This work was performed under contract to the Western Canada Microtox Users Committee. Microtox® test reagent and analytical supplies were supplied by AZUR Environmental, Carlsbad, CA. Statistical analysis was performed by Hai Nguyen and Dr. Zack Florence, ARC Statistical Applications Program.

## **H. REFERENCES**

1. **Baddaloo, E. G.** 1986. Proceeding of Alberta Oil Sands Tailings Wastewater Treatment Technology Workshop. Held at Mildred Lake Research Station, Fort McMurray, Alberta; 1985 October 29-30. Prepared for Alberta Environment, Research Management Division. RMD Report 36/38. 129 pp.
2. **Baker, B. L.** 1985. Surface water contamination by landfill leachates. Prepared for Alberta Environment, Research Management Division by Kananaskis Centre for Environmental Research, University of Calgary. RMD Report L-90. 27 pp.
3. **Beckman.** 1982. Microtox® System Operation Manual #015-555879, Beckman Instruments, Inc., Carlsbad, CA.
4. **Blaise, C., R. Van Coillie, N. Bermingham, and G. Coulombe.** 1987. Comparaison des responses toxiques de trois indicateurs biologiques (bacteries, algues, poissons) exposes a des effluents de fabriques de pates et papiers. *Revue internationale des sciences de l'eau* 3(1): 9-17.
5. **Bulich, A. A., M. W. Greene, and D. L. Isenberg.** 1981. Reliability of the bacterial luminescence assay for determination of the toxicity of pure compounds and complex effluents. in Dr. R. Branson and K. L. Dickson (eds.) *Aquatic Toxicology and Hazard Assessment: Fourth Conference, ASTM STP 737.* American Society for Testing and Materials, Philadelphia, PA pp. 338-347.

6. **Cameron, R. D., E. C. McDonald, M. G. Mager, S. C. Liptak, and P. D. Parkinson.** 1982. Toxicity of landfill leachates. Environ. Canada Rep. EPS 4-EC-82-7. 41 pp.
7. **Coleman, R. N. and A. A. Qureshi.** 1985. Microtox® and *Spirillum volutans* tests for assessing toxicity of environmental samples. Bull. Environ. Contam. Toxicol. 35: 443-451.
8. **Gaudet, I. D.** (ed.) 1994. Standard Procedure for Microtox® analysis. Alberta Environmental Centre, Vegreville, AB. AECV94-G1. 63pp.
9. **MacKinnon, M. D. and J. T. Retallack.** 1982. Preliminary characterization and detoxifications of tails pond water at the Syncrude Canada Ltd. oil sands plant. in P. J. Rand (ed.), Land and Water Issues Related to Energy Development. Ann Arbor Sci. pp. 185-210.
10. **Qureshi, A. A., R. N. Coleman, and J. H. Paran.** 1984. Evaluation and refinement of the Microtox® test for use in toxicity screening. in D. Liu and B. J. Dutka (eds.) Toxicity Screening Procedures Using Bacterial Systems. Marcel Dekker Ind., New York NY. pp. 1-22.
11. **Ribo, J. M.** 1997. Interlaboratory Comparison Studies of the Luminescent Bacteria Toxicity Bioassay. Environ. Toxicol. Water Qual. 12: 283-294.
12. **SAS Institute Inc.** 1990a. *SAS System for Linear Models, Third Edition*, Cary, NC: SAS Institute Inc., 329 pp.
13. **SAS Institute Inc.** 1990b. *SAS/STAT User's Guide, Version 6, Fourth Edition, Volume 2*.
14. **SAS Institute Inc.** 1994. *SAS/STAT Software: Changes and Enhancements, Release 6.10*, Cary, NC: SAS Institute Inc., 121 pp.
15. **Stroscher, M. T.** 1984. A comparison of biological testing methods in association with chemical analyses to evaluate toxicity of waste drilling fluids in Alberta. Volume 1. Canadian Petroleum Association Calgary, AB.
16. **Vasseur, P., J. F. Ferand, C. Rast, and G. Larbaight.** 1984 Luminescent marine bacteria in ecotoxicity screening test of complex effluents. in D. Liu and B. J. Dutka (eds.) Toxicity Screening Procedures Using Bacterial Systems, M. Dekker Inc. New York, NY. pp. 23-35.

## I. APPENDICES

### 1. AZUR Quality Control Data

Phenol Quality Control Chart  
Zinc Sulphate Quality Control Chart  
Certificate of Performance, Lot ACV009-2  
Certificate of Performance, Lot ACV009-3  
Certificate of Performance, Lot ACV010-3

### 2. Tables and Figures

#### a) Zinc

Table 1. Results of Zinc Sulphate Assays  
Table 2. Zinc Sulphate EC<sub>50</sub> 15 minute Data Summary  
Figure 1. Zinc Sulphate EC<sub>50</sub> 15 Minute Data  
Figure 2. Zinc Sulphate EC<sub>50</sub> 15 Minute Data (24 hour data removed)  
Table 3. Zinc Sulphate Differences Among Mean EC<sub>50</sub> Values  
Table 4. Zinc Sulphate Differences Among Means, Reagent Lots EC<sub>50</sub> Values

#### b) Copper

Table 5. Results of Copper Sulphate Assays  
Table 6. Copper Sulphate EC<sub>50</sub> 15 minute Data Summary  
Figure 3. Copper Sulphate EC<sub>50</sub> 15 Minute Data  
Figure 4. Copper Sulphate EC<sub>50</sub> 15 Minute Data (24 hour data removed)  
Figure 5. Copper Sulphate EC<sub>50</sub> 15 Minute Data (8 and 24 hour data removed)  
Table 7. Copper Sulphate Differences Among Mean EC<sub>50</sub> Values  
Table 8. Copper Sulphate Differences Among Means, Reagent Lots EC<sub>50</sub> Values

#### c) Phenol

Table 9. Results of Phenol Assays  
Table 10. Phenol EC<sub>50</sub> 5 minute Data Summary  
Figure 6. Phenol EC<sub>50</sub> 5 minute Data  
Table 11. Phenol EC<sub>50</sub> 15 minute Data Summary  
Figure 7. Phenol EC<sub>50</sub> 15 minute Data  
Table 12. Phenol Differences Among Mean EC<sub>50</sub> Values  
Table 13. Phenol Differences Among Means, Reagent Lots EC<sub>50</sub> Values

#### d) Sodium Dodecyl Sulphate

Table 14. Results of Sodium Dodecyl Sulphate Assays  
Table 15. Sodium Dodecyl Sulphate EC<sub>50</sub> 15 minute Data Summary  
Figure 8. Sodium Dodecyl Sulphate EC<sub>50</sub> 15 minute Data  
Figure 9. Sodium Dodecyl Sulphate EC<sub>50</sub> 15 minute Data (24 hour data removed)  
Table 16. Sodium Dodecyl Sulphate Differences Among Mean EC<sub>50</sub> Values  
Table 17. Sodium Dodecyl Sulphate Differences Among Means, Reagent Lots EC<sub>50</sub> Values

**Phenol Quality Control Chart**

**Zinc Sulphate Quality Control Chart**

**Certificate of Analysis Lot ACV009-2**

**Certificate of Analysis Lot ACV009-3**

**Certificate of Analysis Lot ACV010-3**

**Table 1: Results of Zinc Sulphate Assays**

REAGENT LOT	REPLICATE	TIME (H)	EC <sub>20</sub> 5 MIN (mg/L Zn <sup>+2</sup> )	EC <sub>50</sub> 5 MIN (mg/L Zn <sup>+2</sup> )	EC <sub>20</sub> 15 MIN (mg/L Zn <sup>+2</sup> )	EC <sub>50</sub> 15 MIN (mg/L Zn <sup>+2</sup> )
ACV009-2	1	0.50	3.515	15.843	0.382	1.243
ACV009-2	1	0.75	1.335	12.671	0.995	1.791
ACV009-2	1	1.00	4.969	13.562	0.485	1.452
ACV009-2	1	2.00	3.127	15.607	0.822	2.123
ACV009-2	1	4.00	2.819	15.420	0.734	1.875
ACV009-2	1	8.00	1.624	15.021	0.367	1.204
ACV009-2	1	24.00	0.560	12.513	0.257	0.692
ACV009-2	2	0.50	2.511	11.441	0.379	1.151
ACV009-2	2	0.75	2.336	12.700	0.300	1.106
ACV009-2	2	1.00	5.580	13.844	0.488	1.502
ACV009-2	2	2.00	3.928	13.553	0.599	1.633
ACV009-2	2	4.00	2.506	11.008	0.477	1.377
ACV009-2	2	8.00	1.137	10.537	0.265	1.000
ACV009-2	2	24.00	0.990	3.384	0.163	0.498
ACV009-2	3	0.50	0.373	11.401	0.333	1.108
ACV009-2	3	0.75	2.998	12.933	0.426	1.382
ACV009-2	3	1.00	3.064	12.426	0.551	1.540
ACV009-2	3	2.00	3.252	13.666	0.625	1.711
ACV009-2	3	4.00	2.444	14.017	0.558	1.621
ACV009-2	3	8.00	1.854	11.709	0.362	1.182
ACV009-2	3	24.00	1.054	3.807	0.237	0.640
ACV009-3	1	0.50	4.007	13.043	0.419	1.367
ACV009-3	1	0.75	3.274	14.299	0.294	1.101
ACV009-3	1	1.00	2.924	12.545	0.252	1.584
ACV009-3	1	2.00	3.569	15.121	0.679	1.970
ACV009-3	1	4.00	4.130	13.741	0.421	1.497
ACV009-3	1	8.00	1.709	11.611	0.272	1.027
ACV009-3	1	24.00	1.197	3.466	0.171	0.545
ACV009-3	2	0.50	1.097	11.657	0.243	0.838
ACV009-3	2	0.75	2.915	14.051	0.252	1.015
ACV009-3	2	1.00	4.647	13.963	0.555	1.709
ACV009-3	2	2.00	1.059	17.109	0.338	1.229
ACV009-3	2	4.00	3.382	14.816	0.370	1.278
ACV009-3	2	8.00	1.933	10.646	0.238	0.920
ACV009-3	2	24.00	1.023	2.287	0.211	0.548
ACV009-3	3	0.50	4.304	13.990	0.292	1.099
ACV009-3	3	0.75	3.613	15.836	0.200	0.835
ACV009-3	3	1.00	4.540	13.868	0.676	1.887
ACV009-3	3	2.00	3.643	13.278	0.585	1.653
ACV009-3	3	4.00	2.017	10.803	0.461	1.420
ACV009-3	3	8.00	1.632	11.993	0.214	0.896
ACV009-3	3	24.00	0.976	2.814	0.159	0.469
ACV010-3	1	0.50	4.916	30.314	0.163	1.123
ACV010-3	1	0.75	4.318	16.521	0.561	1.761
ACV010-3	1	1.00	7.225	25.235	0.508	1.893
ACV010-3	1	2.00	3.106	24.445	0.346	1.407
ACV010-3	1	4.00	3.666	39.099	0.393	1.592
ACV010-3	1	8.00	4.048	29.766	0.429	1.607

REAGENT LOT	REPLICATE	TIME (H)	EC <sub>20</sub> 5 MIN (mg/L Zn <sup>+2</sup> )	EC <sub>50</sub> 5 MIN (mg/L Zn <sup>+2</sup> )	EC <sub>20</sub> 15 MIN (mg/L Zn <sup>+2</sup> )	EC <sub>50</sub> 15 MIN (mg/L Zn <sup>+2</sup> )
ACV010-3	1	24.00	0.936	2.698	0.110	0.383
ACV010-3	2	0.50	8.898	33.495	0.373	1.706
ACV010-3	2	0.75	7.000	25.200	0.308	1.494
ACV010-3	2	1.00	7.164	26.968	0.295	1.489
ACV010-3	2	2.00	3.234	35.945	0.239	1.333
ACV010-3	2	4.00	4.344	24.101	0.527	1.883
ACV010-3	2	8.00	3.596	8.863	0.336	1.456
ACV010-3	2	24.00	1.292	3.412	0.158	0.519
ACV010-3	3	0.50	4.370	34.014	0.315	1.625
ACV010-3	3	0.75	5.903	26.135	0.243	1.173
ACV010-3	3	1.00	6.643	22.430	0.370	1.534
ACV010-3	3	2.00	5.636	25.826	0.487	1.754
ACV010-3	3	4.00	3.423	23.910	0.295	1.309
ACV010-3	3	8.00	3.969	43.350	0.264	1.201
ACV010-3	3	24.00	0.571	3.016	0.110	0.368

Table 2: Zinc Sulphate EC<sub>50</sub> 15 minute Data Summary (mg/L Zn<sup>+2</sup>)

REAGENT LOT	REPLICATE	TIME (H)							Mean
		0.5	0.75	1	2	4	8	24	
ACV009-2	1	1.243	1.791	1.452	2.123	1.875	1.204	0.692	1.483
	2	1.151	1.106	1.502	1.633	1.377	1.000	0.498	1.181
	3	1.108	1.382	1.540	1.711	1.621	1.182	0.640	1.312
ACV009-2 Mean		1.167	1.426	1.498	1.822	1.624	1.129	0.610	1.325
ACV009-3	1	1.367	1.101	1.584	1.970	1.497	1.027	0.545	1.299
	2	0.838	1.015	1.709	1.229	1.278	0.920	0.548	1.077
	3	1.099	0.835	1.887	1.653	1.420	0.896	0.469	1.180
ACV009-3 Mean		1.101	0.984	1.727	1.617	1.398	0.948	0.521	1.185
ACV010-3	1	1.123	1.761	1.893	1.407	1.592	1.607	0.383	1.395
	2	1.706	1.494	1.489	1.333	1.883	1.456	0.519	1.411
	3	1.625	1.173	1.534	1.754	1.309	1.201	0.368	1.281
ACV010-3 Mean		1.485	1.476	1.639	1.498	1.595	1.421	0.423	1.362
Overall Mean		1.251	1.295	1.621	1.646	1.539	1.166	0.518	1.291

Figure 1: Zinc Sulphate EC<sub>50</sub> 15 Minute Data





Figure 2: Zinc Sulphate EC<sub>50</sub> 15 Minute Data (24 hour data removed)



**Table 3. Zinc Sulphate Differences Among Mean EC<sub>50</sub> Values (mg/L Zn<sup>+2</sup>)**

Reagent		Time						
		0.5	0.75	1	2	4	8	24
ACV009-2	Arithmetic Mean	1.17	1.43	1.50	1.82	1.62	1.13	<b>0.61*</b>
	Standard Error	0.12	0.12	0.12	0.12	0.12	0.12	0.12
ACV009-3	Arithmetic Mean	<b>1.10*</b>	<b>0.98*</b>	1.73	1.62	1.40	<b>0.95*</b>	<b>0.52*</b>
	Standard Error	0.11	0.11	0.11	0.11	0.11	0.11	0.11
ACV010-3	Arithmetic Mean	1.48	1.48	1.64	1.50	1.59	1.42	<b>0.42*</b>
	Standard Error	0.14	0.14	0.14	0.14	0.14	0.14	0.14

Note: Within each reagent, symbol '\*' indicates that the mean is significantly different from the mean for 1 hour duration at 95% confidence level

**Table 4. Zinc Sulphate Differences Among Means, Reagent Lots EC<sub>50</sub> Values (mg/L Zn<sup>+2</sup>)**

Time		Reagent			Multiple Comparisons*
		ACV009-2	ACV009-3	ACV010-3	
0.50	Adjusted Mean	1.53	1.38	1.61	aaa
	Standard Error	0.08	0.08	0.08	
0.75	Adjusted Mean	1.52	1.37	1.60	aaa
	Standard Error	0.07	0.07	0.07	
1.00	Adjusted Mean	1.51	1.36	1.59	aaa
	Standard Error	0.07	0.07	0.07	
2.00	Adjusted Mean	1.47	1.32	1.54	aaa
	Standard Error	0.07	0.07	0.07	
4.00	Adjusted Mean	1.39	1.25	1.45	aaa
	Standard Error	0.06	0.06	0.06	
8.00	Adjusted Mean	1.24	1.10	1.26	aaa
	Standard Error	0.06	0.06	0.06	
24.00	Adjusted Mean	0.63	0.52	0.50	aaa
	Standard Error	0.16	0.16	0.16	

Note: Within each time point, different letters indicate that the adjusted means are significantly different from each other at the 95% confidence level, i.e. no reagent means at a given time are different when the three letters are the same.

**Table 5: Results of Copper Sulphate Assays**

REAGENT LOT	REPLICATE	TIME (H)	EC <sub>20</sub>	EC <sub>50</sub>	EC <sub>20</sub>	EC <sub>50</sub>
			5 MIN (mg/L Cu <sup>+2</sup> )	5 MIN (mg/L Cu <sup>+2</sup> )	15 MIN (mg/L Cu <sup>+2</sup> )	15 MIN (mg/L Cu <sup>+2</sup> )
ACV009-2	1	0.5	0.205	0.504	0.155	0.272
ACV009-2	1	0.75	0.197	0.407	0.103	0.247
ACV009-2	1	1	0.215	0.353	0.105	0.208
ACV009-2	1	2	0.251	0.545	0.133	0.268
ACV009-2	1	4	0.281	0.478	0.171	0.340
ACV009-2	1	8	0.328	0.576	0.214	0.396
ACV009-2	1	24	0.185	0.433	0.120	0.387
ACV009-2	2	0.5	0.136	0.326	0.100	0.187
ACV009-2	2	0.75	0.202	0.352	0.101	0.204
ACV009-2	2	1	0.146	0.323	0.077	0.180
ACV009-2	2	2	0.184	0.368	0.126	0.231
ACV009-2	2	4	0.177	0.490	0.127	0.265
ACV009-2	2	8	0.269	0.879	0.178	0.358
ACV009-2	2	24	0.403	0.543	0.146	0.316
ACV009-2	3	0.5	0.133	0.253	0.093	0.156
ACV009-2	3	0.75	0.144	0.294	0.095	0.172
ACV009-2	3	1	0.179	0.327	0.096	0.204
ACV009-2	3	2	0.193	0.395	0.117	0.222
ACV009-2	3	4	0.223	0.536	0.101	0.279
ACV009-2	3	8	0.383	0.976	0.208	0.404
ACV009-2	3	24	0.579	1.489	0.284	0.523
ACV009-3	1	0.5	0.268	0.578	0.147	0.237
ACV009-3	1	0.75	0.150	0.744	0.080	0.214
ACV009-3	1	1	0.212	0.603	0.110	0.231
ACV009-3	1	2	0.160	0.593	0.129	0.222
ACV009-3	1	4	0.493	1.034	0.213	0.406
ACV009-3	1	8	0.395	1.317	0.226	0.411
ACV009-3	1	24	0.538	1.955	0.210	0.463
ACV009-3	2	0.5	0.177	0.411	0.078	0.192
ACV009-3	2	0.75	0.241	0.586	0.102	0.244
ACV009-3	2	1	0.201	0.521	0.117	0.249
ACV009-3	2	2	0.183	0.567	0.131	0.271
ACV009-3	2	4	0.239	0.518	0.119	0.278
ACV009-3	2	8	0.509	1.161	0.228	0.495
ACV009-3	2	24	0.578	1.669	0.253	0.496
ACV009-3	3	0.5	0.218	0.505	0.137	0.240
ACV009-3	3	0.75	0.236	0.573	0.152	0.282
ACV009-3	3	1	0.201	0.522	0.169	0.282
ACV009-3	3	2	0.239	0.605	0.141	0.283
ACV009-3	3	4	0.242	0.691	0.242	0.356
ACV009-3	3	8	0.596	0.797	0.243	0.456
ACV009-3	3	24	0.622	1.980	0.221	0.572
ACV010-3	1	0.5	0.220	0.534	0.115	0.268
ACV010-3	1	0.75	0.251	0.640	0.126	0.290

REAGENT LOT	REPLICATE	TIME (H)	EC <sub>20</sub>	EC <sub>50</sub>	EC <sub>20</sub>	EC <sub>50</sub>
			5 MIN (mg/L Cu <sup>+2</sup> )	5 MIN (mg/L Cu <sup>+2</sup> )	15 MIN (mg/L Cu <sup>+2</sup> )	15 MIN (mg/L Cu <sup>+2</sup> )
ACV010-3	1	1	0.318	0.622	0.149	0.313
ACV010-3	1	2	0.300	0.523	0.147	0.261
ACV010-3	1	4	0.325	0.949	0.180	0.320
ACV010-3	1	8	0.401	0.924	0.160	0.385
ACV010-3	1	24	0.532	1.010	0.249	0.469
ACV010-3	2	0.5	0.221	0.480	0.120	0.224
ACV010-3	2	0.75	0.385	0.654	0.192	0.331
ACV010-3	2	1	0.290	0.558	0.120	0.245
ACV010-3	2	2	0.231	0.447	0.112	0.246
ACV010-3	2	4	0.258	0.463	0.153	0.290
ACV010-3	2	8	0.332	0.557	0.214	0.378
ACV010-3	2	24	0.444	1.077	0.231	0.481
ACV010-3	3	0.5	0.249	0.449	0.135	0.240
ACV010-3	3	0.75	0.263	0.828	0.145	0.230
ACV010-3	3	1	0.288	0.574	0.125	0.240
ACV010-3	3	2	0.194	0.664	0.096	0.216
ACV010-3	3	4	0.336	1.049	0.201	0.374
ACV010-3	3	8	0.627	1.207	0.233	0.473
ACV010-3	3	24	0.556	2.859	0.200	0.446

**Table 6: Copper Sulphate EC<sub>50</sub> 15 minute Data Summary (mg/L Cu<sup>+2</sup>)**

REAGENT LOT	REPLICATE	TIME (H)							Mean
		0.5	0.75	1	2	4	8	24	
ACV009-2	1	0.272	0.247	0.208	0.268	0.340	0.396	0.387	0.303
	2	0.187	0.204	0.180	0.231	0.265	0.358	0.316	0.249
	3	0.156	0.172	0.204	0.222	0.279	0.404	0.523	0.280
ACV009-2 Mean		0.205	0.208	0.197	0.240	0.295	0.386	0.409	0.277
ACV009-3	1	0.237	0.214	0.231	0.222	0.406	0.411	0.463	0.312
	2	0.192	0.244	0.249	0.271	0.278	0.495	0.496	0.318
	3	0.240	0.282	0.282	0.283	0.356	0.456	0.572	0.353
ACV009-3 Mean		0.223	0.247	0.254	0.259	0.347	0.454	0.510	0.328
ACV010-3	1	0.268	0.290	0.313	0.261	0.320	0.385	0.469	0.329
	2	0.224	0.331	0.245	0.246	0.290	0.378	0.481	0.314
	3	0.240	0.230	0.240	0.216	0.374	0.473	0.446	0.317
ACV010-3 Mean		0.244	0.284	0.266	0.241	0.328	0.412	0.465	0.320
Overall Mean		0.224	0.246	0.239	0.247	0.323	0.417	0.461	0.308

Figure 3: Copper Sulphate 15 Minute EC<sub>50</sub> Data




Figure 4: Copper Sulphate 15 Minute EC<sub>50</sub> Data (24 Hour Data Removed)





Figure 5: Copper Sulphate 15 Minute EC<sub>50</sub> Data (8 and 24 Hour Data Removed)



**Table 7. Copper Sulphate Differences Among Mean EC<sub>50</sub> Values (mg/L Cu<sup>+2</sup>)**

Reagent		Time						
		0.5	0.75	1	2	4	8	24
ACV009-2	Arithmetic Mean	0.21	0.21	0.20	0.24	0.29	<b>0.38*</b>	<b>0.41*</b>
	Standard Error	0.03	0.03	0.03	0.03	0.03	0.03	0.03
ACV009-3	Arithmetic Mean	0.22	0.25	0.25	0.26	0.35	<b>0.45*</b>	<b>0.51*</b>
	Standard Error	0.02	0.02	0.02	0.02	0.02	0.02	0.02
ACV010-3	Arithmetic Mean	0.24	0.28	0.27	0.24	0.33	<b>0.41*</b>	<b>0.47*</b>
	Standard Error	0.02	0.02	0.02	0.02	0.02	0.02	0.02

Note: Within each reagent, symbol '\*' indicates that the mean is significantly different from the mean for 1 hour duration at 95% confidence level

**Table 8. Copper Sulphate Differences Among Means, Reagent Lots EC<sub>50</sub> Values (mg/L Cu<sup>+2</sup>)**

Time		Reagent			Multiple Comparisons*
		ACV009-2	ACV009-3	ACV010-3	
0.50	Adjusted Mean	0.23	0.26	0.27	aaa
	Standard Error	0.02	0.02	0.02	
0.75	Adjusted Mean	0.23	0.27	0.27	aaa
	Standard Error	0.02	0.02	0.02	
1.00	Adjusted Mean	0.23	0.27	0.28	aaa
	Standard Error	0.01	0.01	0.01	
2.00	Adjusted Mean	0.24	0.28	0.29	aaa
	Standard Error	0.01	0.01	0.01	
4.00	Adjusted Mean	0.26	0.31	0.30	<b>aba</b>
	Standard Error	0.01	0.01	0.01	
8.00	Adjusted Mean	0.30	0.35	0.34	<b>aba</b>
	Standard Error	0.01	0.01	0.01	
24.00	Adjusted Mean	0.44	0.55	0.49	aaa
	Standard Error	0.03	0.03	0.03	

Note: Within each time point, different letters indicate that the adjusted means are significantly different from each other at the 95% confidence level, i.e. no reagent means at a given time are different when the three letters are the same.

**Table 9: Results of Phenol Assays**

REAGENT LOT	REPLICATE	TIME (H)	EC <sub>20</sub> 5 MIN (mg/L PHENOL)	EC <sub>50</sub> 5 MIN (mg/L PHENOL)	EC <sub>20</sub> 15 MIN (mg/L PHENOL)	EC <sub>50</sub> 15 MIN (mg/L PHENOL)
ACV009-2	1	0.5	4.398	22.150	5.220	24.417
ACV009-2	1	0.75	5.010	22.808	6.346	25.989
ACV009-2	1	1	5.303	22.816	5.933	24.730
ACV009-2	1	2	3.902	18.935	4.295	19.489
ACV009-2	1	4	4.675	20.564	5.812	22.494
ACV009-2	1	8	5.595	23.355	5.413	22.029
ACV009-2	1	24	4.256	17.978	4.467	19.529
ACV009-2	2	0.5	5.793	24.571	6.584	27.061
ACV009-2	2	0.75	4.614	20.685	6.241	24.596
ACV009-2	2	1	4.675	21.320	5.320	22.753
ACV009-2	2	2	3.946	19.305	4.460	20.580
ACV009-2	2	4	5.376	22.476	5.267	23.085
ACV009-2	2	8	5.931	25.228	6.306	24.457
ACV009-2	2	24	5.533	20.300	5.443	20.984
ACV009-2	3	0.5	4.268	22.115	5.950	26.076
ACV009-2	3	0.75	5.942	23.697	6.523	25.556
ACV009-2	3	1	5.448	23.056	7.561	26.862
ACV009-2	3	2	4.561	20.091	4.986	21.124
ACV009-2	3	4	4.122	20.027	4.670	20.900
ACV009-2	3	8	5.430	22.395	5.509	21.977
ACV009-2	3	24	4.024	16.966	4.977	20.895
ACV009-3	1	0.5	3.727	19.493	4.717	22.661
ACV009-3	1	0.75	4.858	20.859	4.552	19.846
ACV009-3	1	1	4.016	18.207	5.046	21.801
ACV009-3	1	2	3.966	17.545	4.457	19.675
ACV009-3	1	4	4.032	17.557	4.012	18.118
ACV009-3	1	8	5.050	20.482	4.523	19.522
ACV009-3	1	24	4.496	18.027	4.374	18.419
ACV009-3	2	0.5	3.609	18.326	4.363	19.968
ACV009-3	2	0.75	4.781	19.938	5.314	22.130
ACV009-3	2	1	5.608	22.130	6.069	22.991
ACV009-3	2	2	4.609	19.080	5.337	20.614
ACV009-3	2	4	4.168	19.364	4.754	19.858
ACV009-3	2	8	4.592	20.539	4.946	21.736
ACV009-3	2	24	4.999	20.356	6.122	22.539
ACV009-3	3	0.5	6.060	23.856	5.079	22.119
ACV009-3	3	0.75	4.874	22.871	5.173	23.098
ACV009-3	3	1	5.359	21.983	5.206	21.766
ACV009-3	3	2	4.284	18.682	5.116	20.487
ACV009-3	3	4	4.389	19.629	4.538	20.119
ACV009-3	3	8	5.706	21.415	6.202	22.906
ACV009-3	3	24	4.662	19.102	5.109	20.703
ACV010-3	1	0.5	3.467	17.621	3.668	18.459
ACV010-3	1	0.75	4.802	20.991	5.772	23.385


Western Canada Microtox® Users Committee – Final Report

REAGENT LOT	REPLICATE	TIME (H)	EC <sub>20</sub> 5 MIN (mg/L PHENOL)	EC <sub>50</sub> 5 MIN (mg/L PHENOL)	EC <sub>20</sub> 15 MIN (mg/L PHENOL)	EC <sub>50</sub> 15 MIN (mg/L PHENOL)
ACV010-3	1	1	3.785	18.560	4.851	20.865
ACV010-3	1	2	4.204	17.629	3.973	17.271
ACV010-3	1	4	4.344	16.924	4.860	17.729
ACV010-3	1	8	4.877	19.829	4.336	18.423
ACV010-3	1	24	4.089	19.003	4.114	20.097
ACV010-3	2	0.5	4.743	20.862	4.959	21.464
ACV010-3	2	0.75	5.305	21.180	6.670	24.600
ACV010-3	2	1	5.669	22.074	5.324	21.526
ACV010-3	2	2	3.325	15.299	3.834	16.601
ACV010-3	2	4	4.172	18.411	4.183	17.880
ACV010-3	2	8	5.368	21.869	5.264	21.228
ACV010-3	2	24	4.310	18.775	4.595	19.651
ACV010-3	3	0.5	4.985	22.084	5.078	22.851
ACV010-3	3	0.75	5.582	22.353	5.926	23.313
ACV010-3	3	1	4.173	18.125	4.666	19.353
ACV010-3	3	2	3.626	16.390	3.519	16.544
ACV010-3	3	4	3.890	17.516	4.334	18.666
ACV010-3	3	8	5.298	20.849	5.415	21.594
ACV010-3	3	24	3.730	15.916	4.213	17.878

**Table 10: Phenol EC<sub>50</sub> 5 minute Data Summary (mg/L)**

REAGENT LOT	REPLICATE	TIME (H)							Mean
		0.5	0.75	1	2	4	8	24	
ACV009-2	1	22.15	22.808	22.816	18.935	20.564	23.355	17.978	21.229
	2	24.571	20.685	21.32	19.305	22.476	25.228	20.2999	21.984
	3	22.115	23.697	23.056	20.091	20.027	22.395	16.966	21.192
ACV009-2 Mean		22.945	22.397	22.397	19.444	21.022	23.659	18.415	21.468
ACV009-3	1	19.493	20.859	18.207	17.545	17.557	20.482	18.027	18.881
	2	18.326	19.938	22.13	19.08	19.364	20.539	20.356	19.962
	3	23.856	22.871	21.983	18.682	19.629	21.415	19.102	21.077
ACV009-3 Mean		20.558	21.223	20.773	18.436	18.850	20.812	19.162	19.973
ACV010-3	1	17.621	20.991	18.56	17.629	16.924	19.829	19.003	18.651
	2	20.862	21.18	22.074	15.299	18.411	21.869	18.775	19.781
	3	22.084	22.353	18.125	16.39	17.516	20.849	15.916	19.033
ACV010-3 Mean		20.189	21.508	19.586	16.439	17.617	20.849	17.898	19.155
Overall Mean		21.231	21.709	20.919	18.106	19.163	21.773	18.491	20.199


Figure 6: Phenol EC<sub>50</sub> 5 minute Data



**Table 11: Phenol EC<sub>50</sub> 15 minute Data Summary (mg/L)**

REAGENT LOT	REPLICATE	TIME (H)							Mean
		0.5	0.75	1	2	4	8	24	
ACV009-2	1	24.417	25.989	24.730	19.489	22.494	22.029	19.529	22.668
	2	27.061	24.596	22.753	20.580	23.085	24.457	20.984	23.359
	3	26.076	25.556	26.862	21.124	20.900	21.977	20.895	23.341
ACV009-2 Mean		25.851	25.380	24.782	20.398	22.160	22.821	20.469	23.123
ACV009-3	1	22.661	19.846	21.801	19.675	18.118	19.522	18.419	20.006
	2	19.968	22.130	22.991	20.614	19.858	21.736	22.539	21.405
	3	22.119	23.098	21.766	20.487	20.119	22.906	20.703	21.600
ACV009-3 Mean		21.583	21.691	22.186	20.259	19.365	21.388	20.554	21.004
ACV010-3	1	18.459	23.385	20.865	17.271	17.729	18.423	20.097	19.461
	2	21.464	24.600	21.526	16.601	17.880	21.228	19.651	20.421
	3	22.851	23.313	19.353	16.544	18.666	21.594	17.878	20.028
ACV010-3 Mean		20.925	23.766	20.581	16.805	18.092	20.415	19.209	19.970
Overall Mean		22.786	23.613	22.516	19.154	19.872	21.541	20.077	21.366

Figure 7: Phenol EC<sub>50</sub> 15 minute Data



**Table 12. Phenol Differences Among Mean EC<sub>50</sub> Values (mg/L)**

Reagent		Time						
		0.5	0.75	1	2	4	8	24
ACV009-2	Arithmetic Mean	25.85	25.38	24.78	<b>20.40*</b>	22.16	22.82	<b>20.47*</b>
	Standard Error	0.73	0.73	0.73	0.73	0.73	0.73	0.73
ACV009-3	Arithmetic Mean	21.58	21.69	22.19	20.26	19.37	21.39	20.55
	Standard Error	0.82	0.82	0.82	0.82	0.82	0.82	0.82
ACV010-3	Arithmetic Mean	20.92	<b>23.77*</b>	20.58	<b>16.81*</b>	18.09	20.42	19.21
	Standard Error	0.74	0.74	0.74	0.74	0.74	0.74	0.74

Note: Within each reagent, symbol '\*' indicates that the mean is significantly different from the mean for 1 hour duration at 95% confidence level

**Table 13. Phenol Differences Among Means, Reagent Lots EC<sub>50</sub> Values (mg/L)**

Time		Reagent			Multiple Comparisons*
		ACV009-2	ACV009-3	ACV010-3	
0.50	Adjusted mean	23.99	21.17	20.27	<b>abb</b>
	Standard error	0.53	0.53	0.53	
0.75	Adjusted mean	23.94	21.16	20.26	<b>abb</b>
	Standard error	0.52	0.52	0.52	
1.00	Adjusted mean	23.90	21.16	20.24	<b>abb</b>
	Standard error	0.52	0.52	0.52	
2.00	Adjusted mean	23.74	21.12	20.19	<b>abb</b>
	Standard error	0.49	0.49	0.49	
4.00	Adjusted mean	23.41	21.06	20.07	<b>abb</b>
	Standard error	0.45	0.45	0.45	
8.00	Adjusted mean	22.75	20.93	19.84	<b>abb</b>
	Standard error	0.46	0.46	0.46	
24.00	Adjusted mean	20.13	20.42	18.93	aaa
	Standard error	1.12	1.12	1.12	

Note: Within each time point, different letters indicate that the adjusted means are significantly different from each other at the 95% confidence level, i.e. no reagent means at a given time are different when the three letters are the same.

Table 14: Results of Sodium Dodecyl Sulphate Assays

REAGENT LOT	REPLICATE	TIME (H)	EC <sub>20</sub>	EC <sub>50</sub>	EC <sub>20</sub>	EC <sub>50</sub>
			5 MIN (mg/L SDS)	5 MIN (mg/L SDS)	15 MIN (mg/L SDS)	15 MIN (mg/L SDS)
ACV009-2	1	0.5	0.738	1.855	0.372	0.855
ACV009-2	1	0.75	1.180	2.038	0.480	0.953
ACV009-2	1	1	0.597	1.575	0.290	0.736
ACV009-2	1	2	0.598	1.466	0.269	0.654
ACV009-2	1	4	0.859	1.757	0.287	0.721
ACV009-2	1	8	0.608	1.762	0.226	0.632
ACV009-2	1	24	0.897	2.544	0.332	0.973
ACV009-2	2	0.5	0.840	2.140	0.352	0.853
ACV009-2	2	0.75	0.811	1.872	0.356	0.821
ACV009-2	2	1	0.845	1.879	0.366	0.845
ACV009-2	2	2	0.467	1.344	0.237	0.603
ACV009-2	2	4	0.542	1.421	0.197	0.534
ACV009-2	2	8	0.620	1.522	0.220	0.621
ACV009-2	2	24	0.928	2.451	0.382	1.095
ACV009-2	3	0.5	0.962	1.917	0.418	0.920
ACV009-2	3	0.75	0.890	1.838	0.402	0.873
ACV009-2	3	1	0.645	1.784	0.301	0.770
ACV009-2	3	2	0.583	1.442	0.239	0.598
ACV009-2	3	4	0.511	1.364	0.187	0.503
ACV009-2	3	8	0.500	1.329	0.167	0.465
ACV009-2	3	24	1.172	2.697	0.559	1.359
ACV009-3	1	0.5	1.035	1.961	0.293	0.882
ACV009-3	1	0.75	1.026	1.996	0.391	0.979
ACV009-3	1	1	0.945	1.861	0.395	0.862
ACV009-3	1	2	0.872	1.783	0.365	0.805
ACV009-3	1	4	0.578	1.597	0.220	0.593
ACV009-3	1	8	0.423	1.528	0.164	0.523
ACV009-3	1	24	1.057	2.523	0.365	1.174
ACV009-3	2	0.5	1.178	1.974	0.498	1.003
ACV009-3	2	0.75	0.809	1.598	0.409	0.852
ACV009-3	2	1	0.971	1.892	0.415	0.900
ACV009-3	2	2	0.693	1.620	0.368	0.801
ACV009-3	2	4	0.576	1.503	0.230	0.629
ACV009-3	2	8	0.629	1.686	0.207	0.625
ACV009-3	2	24	0.897	2.484	0.302	1.001
ACV009-3	3	0.5	0.827	1.673	0.346	0.818
ACV009-3	3	0.75	0.765	1.605	0.381	0.826
ACV009-3	3	1	0.764	1.672	0.321	0.770
ACV009-3	3	2	0.590	1.477	0.271	0.668
ACV009-3	3	4	0.597	1.569	0.251	0.674
ACV009-3	3	8	0.650	1.811	0.242	0.691
ACV009-3	3	24	1.502	5.722	0.594	1.913
ACV010-3	1	0.5	0.823	1.817	0.380	0.860
ACV010-3	1	0.75	0.832	1.924	0.361	0.895

REAGENT LOT	REPLICATE	TIME (H)	EC <sub>20</sub>	EC <sub>50</sub>	EC <sub>20</sub>	EC <sub>50</sub>
			5 MIN (mg/L SDS)	5 MIN (mg/L SDS)	15 MIN (mg/L SDS)	15 MIN (mg/L SDS)
ACV010-3	1	1	0.798	1.743	0.378	0.854
ACV010-3	1	2	0.775	1.615	0.417	0.848
ACV010-3	1	4	0.601	1.419	0.333	0.727
ACV010-3	1	8	0.704	1.345	0.487	0.819
ACV010-3	1	24	2.737	12.471	0.942	3.024
ACV010-3	2	0.5	0.743	1.815	0.329	0.806
ACV010-3	2	0.75	0.841	1.881	0.370	0.851
ACV010-3	2	1	0.844	1.700	0.415	0.899
ACV010-3	2	2	0.853	1.622	0.475	0.915
ACV010-3	2	4	0.672	1.389	0.258	0.636
ACV010-3	2	8	0.924	2.090	0.327	0.860
ACV010-3	2	24	1.303	5.159	0.658	2.268
ACV010-3	3	0.5	0.693	1.606	0.360	0.782
ACV010-3	3	0.75	0.651	1.340	0.389	0.794
ACV010-3	3	1	0.633	1.554	0.263	0.710
ACV010-3	3	2	0.527	1.305	0.231	0.584
ACV010-3	3	4	0.600	1.468	0.231	0.573
ACV010-3	3	8	0.539	1.838	0.187	0.610
ACV010-3	3	24	1.794	8.864	0.622	2.562

**Table 15: Sodium Dodecyl Sulphate EC<sub>50</sub> 15 minute Data Summary (mg/L)**

REAGENT	REPLICATE	TIME (H)							Mean
		0.5	0.75	1	2	4	8	24	
ACV009-2	1	0.855	0.953	0.736	0.654	0.721	0.632	0.973	0.789
	2	0.853	0.821	0.845	0.603	0.534	0.621	1.095	0.767
	3	0.920	0.873	0.770	0.598	0.503	0.465	1.359	0.784
ACV009-2 Mean		0.876	0.882	0.784	0.618	0.586	0.573	1.142	0.780
ACV009-3	1	0.882	0.979	0.862	0.805	0.593	0.523	1.174	0.831
	2	1.003	0.852	0.900	0.801	0.629	0.625	1.001	0.830
	3	0.818	0.826	0.770	0.668	0.674	0.691	1.913	0.909
ACV009-3 Mean		0.901	0.886	0.844	0.758	0.632	0.613	1.363	0.857
ACV010-3	1	0.860	0.895	0.854	0.848	0.727	0.819	3.024	1.147
	2	0.806	0.851	0.899	0.915	0.636	0.860	2.268	1.034
	3	0.782	0.794	0.710	0.584	0.573	0.610	2.562	0.945
ACV010-3 Mean		0.816	0.847	0.821	0.782	0.645	0.763	2.618	1.042
Overall Mean		0.864	0.872	0.816	0.720	0.621	0.650	1.708	0.893

Figure 8: Sodium Dodecyl Sulphate EC<sub>50</sub> 15 minute Data





Figure 9: Sodium Dodecyl Sulphate EC<sub>50</sub> 15 minute Data (24 hour data removed)



**Table 16. Sodium Dodecyl Sulphate Differences Among Mean EC<sub>50</sub> Values (mg/L)**

Reagent		Time						
		0.5	0.75	1	2	4	8	24
ACV009-2	Arithmetic mean	0.88	0.88	0.78	0.62	0.59	0.57	<b>1.14*</b>
	Standard error	0.06	0.06	0.06	0.06	0.06	0.06	0.06
ACV009-3	Arithmetic mean	0.90	0.89	0.84	0.76	0.63	0.61	<b>1.36*</b>
	Standard error	0.11	0.11	0.11	0.11	0.11	0.11	0.11
ACV010-3	Arithmetic mean	0.82	0.85	0.82	0.78	0.65	0.76	<b>2.62*</b>
	Standard error	0.10	0.10	0.10	0.10	0.10	0.10	0.10

Note: Within each reagent, symbol '\*' indicates that the mean is significantly different from the mean for 1 hour duration at 95% confidence level

**Table 16. Sodium Dodecyl Sulphate Differences Among Means, Reagent Lots EC<sub>50</sub> Values (mg/L)**

Time		Reagent			Multiple Comparisons*
		ACV009-2	ACV009-3	ACV010-3	
0.50	Adjusted mean	0.71	0.75	0.64	aaa
	Standard error	0.06	0.06	0.06	
0.75	Adjusted mean	0.71	0.75	0.66	aaa
	Standard error	0.06	0.06	0.06	
1.00	Adjusted mean	0.71	0.76	0.68	aaa
	Standard error	0.06	0.06	0.06	
2.00	Adjusted mean	0.73	0.78	0.75	aaa
	Standard error	0.06	0.06	0.06	
4.00	Adjusted mean	0.76	0.82	0.91	aaa
	Standard error	0.05	0.05	0.05	
8.00	Adjusted mean	0.81	0.90	1.21	<b>aab</b>
	Standard error	0.06	0.06	0.06	
24.00	Adjusted mean	1.04	1.24	2.44	<b>aab</b>
	Standard error	0.13	0.13	0.13	

Note: Within each time point, different letters indicate that the adjusted means are significantly different from each other at the 95% confidence level, i.e. no reagent means at a given time are different when the three letters are the same.